

High Harmonic Generation as an XUV source for probing Ultrafast Dynamics in Perovskitic Semiconductors

Stavroula Vovla^{1,2}, G. Crippa³, D. Faccialà², L. Poletto², F. Frassetto², A. G. Ciriolo², R. Martínez Vázquez², R. Osellame², S. Stagira¹, E. Cinquanta², C. Vozzi², M. Devetta²

stavroula.vovla@gmail.com

¹Physics Department, Politecnico di Milano, Italy

²Institute of Photonics and Nanotechnologies, Consiglio Nazionale delle Ricerche, Italy

³CEA, Université Paris-Saclay, Gif-sur-Yvette Cedex, France

With conventional semiconductors facing a shortage, the need for alternative resources is evident. One potential candidate is 3D perovskites whose interesting properties – e.g., low exciton binding energy [1], defect tolerance [2], large polarons [3] - constitutes them perfect for light harvesting devices [4].

Condensed matter phenomena that may hinder their performance, like polaron formation, are observed at femtosecond timescales. Ultrafast soft X-ray spectroscopy allows studying this light-matter interaction with an unprecedented temporal and spatial resolution, with the further advantage of being element-selective, oxidation, and spin-state specific. As such, we plan to investigate large polaron formation dynamics in 3D perovskites with time-resolved Near Edge X-ray Absorption Fine Structure (NEXAFS). Pump-probe spectroscopy is one of the most common experimental schemes for these types of measurements. In summary, a first IR pump pulse excites the sample, then the evolution is probed by a second time-delayed XUV pulse. Hence, innovative experimental set-ups that focus on broadband detection, high dynamic range, and sensitivity have become necessary in ultrafast dynamics.

High-order harmonic generation of femtosecond laser pulses interacting with noble gases has become a popular table-top source of broadband coherent XUV radiation throughout the years. By implementing a microfluidic device instead of the trusted gas-jet configuration, we obtained an extended energy cutoff, and a higher generation yield [5, 6]. Furthermore, an integrated solution for filtering IR radiation under vacuum has been investigated. High vacuum systems were also realized to manipulate, detect, and characterize the generated XUV light.

However, all-optical techniques have huge untapped potential, given that conventional spectrometers are characterized by limited spectral ranges (either low or high-energy photons), and a lack of polarization characterization which excludes the studies of magnetic materials and biologically relevant molecules. Therefore, an innovative detection scheme was realized in our laboratory to surmount these limitations.

References

- [1] Manser J. S. et al, Chemical Reviews **116**, 21 (2016).
- [2] Steirer K. X. et al, ACS Energy Letters **1**, 2 (2016).
- [3] Canelli O. et al, J. Am. Chem. Soc. **143**, 15 (2021).
- [4] Karuturi S.K. et al, ACS Appl Mater Interfaces **10**, 28 (2018).
- [5] Ciriolo A. G. et al, Journal of Physics: Photonics **2**, 024005 (2020).
- [6] Ciriolo A. G. et al, APL Photonics **7**, 11 (2022).